

# MTN/2200I-4P Series

Intrinsically safe industrial accelerometer ATEX & IECEx Group II approved

ATEX and IECEx Group II certified. General purpose top-entry constant current accelerometer with isolated AC output. Made from robust stainless steel throughout for long term vibration analysis in harsh, hazardous gas and dust environments. Internal electronics are isolated to minimise noise with increased bias voltage stability. Sealed to IP67 and includes 4-pin M12 connector and is available with a wide range of mountings.


#### **Applications**


- General industry
- Compressors, pumps etc
- Oil and petrochemical

MTN/2200I-4P



#### **Dimensions**





-							
- 1	ρ	Ch	٦ı	าเ	C	ิล	
	-	<b>U</b> I			~	9	

Technical				
Standard sensitivity	100mV/g ±10% nominal @ 80Hz			
Frequency response	2Hz to 10kHz ±5% (-3dB @ 0.8Hz)			
Mounted base resonance	18kHz (nominal)			
Isolation	Base isolated			
Transverse sensitivity	Less than 5%			
Electrical noise	0.1mg max			
Current range	0.5 to 8mA			
Temperature range for	T4 (-55°C ≤ Ta ≤ +115°C) T6 (-55°C ≤ Ta ≤ +65°C)			
Bias voltage	12V DC (nominal)			
Case material	Stainless steel			
Maximum cable length	See system drawing ATX037			
Mounting torque	8Nm			
Weight	170g (nominal)			
Sealing	IP67			
Insulation	Units will pass a 500V insulation test			

#### **Certificate details**

Group II <sup>1</sup>	BAS02ATEX1057X and IECEx BAS 08.0013X Ex II 1GD T135°C Ex ia IIC T4 Ex iaD 20 T135°C (-55°C $\leq$ Ta $\leq$ +115°C)			
	Ex II 1GD T85°C Ex ia IIC T6 Ex iaD 20 T85°C (-55°C $\leq$ Ta $\leq$ +65°C)			
Terminal parameters	Ui = 28V, li = 93mA, Pi = 0.65W For Ci and Li see certificate			
Barrier	1 x MTL7728+ (BAS01ATEX7217) or (P&F Z728 BAS01ATEX7005) or any other barrier that conforms to note 5 of ATX037			



Monitran Ltd | Monitor House | 33 Hazlemere Road | Penn | Bucks | UK | HP10 8AD

Telephone +44 (0)1494 816569 | E-mail info@monitran.com | Website www.monitran.com



## MTN/2200I-4P Series

Intrinsically safe industrial accelerometer ATEX & IECEx Group II approved

Studs and grub screws			Mounting adapters			
Part #	From	То	Part #	From	То	
MS036	¼"-28 UNF Male	M6 Male	MS005	Q/F Male	¼"-28 UNF Female	
MS039	¼"-28 UNF Male	10-32 UNF Male	MS007	Q/F Male	10-32 UNF Female	
MS067	¼"-28 UNF Male	M8 Male	MS008	Q/F Male	M8 Female	
MS068	¼"-28 UNF Male	¼"-28 UNF Male	MS011	¼"-28 UNF Male	Q/F Female	
MS124	¼"-28 UNF Male	M10 Male	MS013	¼"-28 UNF Male	Glue base	
MS132	¼"-28 UNF Male	M12 Male	MS033	¼"-28 UNF Male	Q/F Female	
	I	I	MS038	Q/F Male	M8 Conical Male	
			MS061	¼"-28 UNF Male	10-32 UNF Male	
Quick fit ada	pters		MS079	¼"-28 UNF Male	Q/F Female	
			MS106	Q/F Male	M10 Female	
Part #	From	То	Isolation			
MS001	Q/F Male	Glue base	1301811011		<u> </u>	
MS002	Q/F Male	M8 Male	Part #	From	То	
MS003	Q/F Male	M10 Male	MS034	¼"-28 UNF Male	¼"-28 UNF Female	
MS004	Q/F Male	¼"-28 UNF Male	MS093	Q/F Male	M8 Male	
MS006	Q/F Male	M6 Male				
Options			System conn	ection		
Dust o	g connectors ption (Group II only) sensitivities (see below)	)	Pin 2 Pin 4	Constant c	Constant current diode (0.5 to 8.0mA)	
Part #	Mounting	xx = Optional sensitivity (mV/g) 10		Safety barrier system see drawing ATX037	AC output	
NATNI /22001		10			• ov	

30 50 500

1000

### Note: Care should be taken not to install this in a high velocity dust laden atmosphere.

<sup>1</sup> Warning ref Group II: The Ci and Li were previously lower. The installer must take account of the increase in internal capacitance and inductance present on this apparatus.

Monitran Ltd | Monitor House | 33 Hazlemere Road | Penn | Bucks | UK | HP10 8AD

¼" UNF Female

Q/F Female

Telephone +44 (0)1494 816569 | E-mail info@monitran.com | Website www.monitran.com

MTN/2200I-4P-xx

MTN/2200IQ-4P-xx

<u>sira</u>